

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JUL 13 2010

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130130</u> ✓	
Name <u>JEREMY CONKLIN</u>		ID# <u>6715</u> ✓	Date <u>07.07.10</u> ✓
A	Agency <u>VALDEZ POLICE</u>		Phone # <u>835-4560</u>
Instrument Location <u>UPD JML</u>			
Alco S/N <u>X172988</u> ✓		Target Value <u>.085</u> ✓	High Pressure <u>800</u>
B	Alco Test Value Average <u>.084</u> ✓		
	1 st Alco	2 nd Alco	
Signature <u>[Signature]</u> 271		<u>CAB</u> 7/15/10	
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

(Do Not Write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 27th day of July, 2010

Carolyn M. Noland
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130130

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

JULY 07, 2010

OPERATOR'S NAME:
CONKLING/JEREMY/WHEELER
OPERATOR'S NUMBER: 6715
SUBJECT'S LAST NAME:

VOC
SUBJECT'S FIRST NAME/MI :
VOC

O.L. #: VOC
DEPT/AGENCY: UDZ1
CASE/REPORT: VOC
TEST TYPE: U
ALCO TARGET VALUE: .085
ALCO S/N: X172988

--- BREATH ANALYSIS ---

.085 ADJUSTED FOR 30.43 in
ALCO TARGET .086 09:31
BLANK TEST .000 09:32
INTERNAL STANDARD VERIFIED 09:32
ALCO TV 30.43 in .084 09:32
BLANK TEST .000 09:33
SUBJECT SAMPLE .000 09:34
BLANK TEST .000 09:34
ALCO TV 30.43 in .084 09:35
BLANK TEST .000 09:36

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130130

JULY 07, 2010
TIME 09:29

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 30.43 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢